

LICENSE REQUIREMENTS FOR TEMPERATION INSIDE THERMAL REACTION (EPL POINT 2)

Data reported in summary at end of reporting period

M2 Requirement to monitor concentration of pollutants discharged

M2.1 For each monitoring/discharge point or utilisation area specified below (by a point number), the licensee must monitor (by sampling and obtaining results by analysis) the concentration of each pollutant specified in Column 1. The licensee must use the sampling method, units of measure, and sample at the frequency, specified opposite in the other columns:

M2.2 Air Monitoring Requirements

POINT 1

Pollutant	Units of measure	Frequency	Sampling Method
Hydrogen fluoride	milligrams per cubic metre	Yearly	TM-9
PM10	milligrams per cubic metre	Yearly	OM-5
Sulfur Oxides	milligrams per cubic metre	Yearly	TM-4
Total Solid Particles	milligrams per cubic metre	Yearly	TM-15

POINT 2

Pollutant	Units of measure	Frequency	Sampling Method
Temperature	degrees Celsius	Continuous	TM-2